


# EMBOLIZATION COIL





Stainless steel spiral with polyester blisters and radiopaque markers at the ends.  
For use with 0.035" or 0.038" catheters

Release

THREAD (INCH)	LENGTH (CM)	DIAMETER (MM)	RELEASE
0.032"	2	3	Not controlled
0.032" - 0.035"	4	4	Not controlled
0.032"	5	6	Not controlled
0.032"	5	8	Not controlled
0.032" - 0.035"	8	5	Not controlled
0.032" - 0.035"	8	8	Not controlled
0.032" - 0.035"	10	8	Not controlled
0.032" - 0.035"	10	10	Not controlled
0.032" - 0.035"	15	8	Not controlled
0.032" - 0.035"	15	10	Not controlled
0.032" - 0.035"	15	15	Not controlled

Code Catalog 60.4513  
Rev. 02

Braile Biomédica Indústria, Comércio e Representações Ltda.  
Av. Pres. Juscelino K. de Oliveira, 1505 • Jardim Tarraf I  
Zip Code 15091-450 • São José do Rio Preto • SP • Brazil  
Phone #55 17 2136 7000 • Fax #55 17 2136 7040  
Federal Tax ID: 52.828.936/0001-09  
Technician in charge: Vladimir D. A. Ramirez • CRF/SP 9010  
ANVISA Reg. nº 10159030053  
www.braile.com.br

